See back for additional applications


Boxer's Fracture Application


Warnings and Instructions: Review carefully, proper application is required

Awarning: This device will not prevent or eliminate risk of injury. Do Not Overtighten. If swelling, pain, skin irritation, or an unusual reaction occurs, discontinue use immediately and consult your medical professional. This device should not be worn by persons with known allergies to neoprene.


4. Finished application.

Care: Hand wash using mild soap. Rinse thoroughly. Air dry only. Do not tumble dry.





1. Apply the brace, wrapping the shortest strap around the little and ring fingers. Make sure to slide hand up so the curves in the brace match the mid-hand crease and the wrist.




3. Secure the wrist strap.

OPTIONAL: Form stay to fit



5. Optional A: If there is pressure on the ulnar styloid, you can adjust the dorsal stay. See No. 6.



7. Finished application with extra bend for ulnar styloid pressure relief.



1185 E Main St., Santa Paula, California 93060 California: 800-221-5465, National: 800-654-3241 International: 805-525-4244, fax: 805-933-2348 U.S. fax: 800-559-5975, www.hely-weber.com



Rev D


2/11


Radial Gutter Application

Warnings and Instructions: Review carefully, proper application is required

Warning: This device will not prevent or eliminate risk of injury. Do Not Overtighten, If swelling, pain, skin irritation, or an unusual reaction occurs, discontinue use immediately and consult your medical professional. This device should not be worn by persons with known allergies to neoprene. Care: Hand wash using mild soap. Rinse thoroughly. Air dry only. Do not tumble dry.


Note: This application uses a right hand brace for a left hand injury. Use a left hand brace for a right hand injury.






1. Apply brace, wrapping shortest strap around index and middle fingers, first.

2. Secure the middle strap.



3. Secure the wrist strap.



5. Finished application.



1185 E Main St., Santa Paula, California 93060 California: 800-221-5465, National: 800-654-3241 International: 805-525-4244, fax: 805-933-2348 U.S. fax: 800-559-5975, www.hely-weber.com








4. Finished application.

U.S. Patent No. 7278980, 7402149, 7455650, 7442177



6. Optional B: Hold the dorsal stay between the index finger and thumb directly over the ulnar styloid. Bend the stay to accommodate the ulnar styloid. This will relieve the pressure.

U.S. Patent No. 7278980, 7402149, 7455650, 7442177

MDSS GmbH, Schiffgraben 41

30175 Hannover, Germany


3848

Metacarpal Fracture Application


3848

Warnings and Instructions: Review carefully, proper application is required


Warning: This device will not prevent or eliminate risk of injury. Do Not Overtighten. If swelling, pain, skin irritation, or an unusual reaction occurs, discontinue use immediately and consult your medical professional. This device should not be worn by persons with known allergies to neoprene. Care: Hand wash using mild soap. Rinse thoroughly. Air dry only. Do not tumble dry.




 Apply brace by putting small finger through hole. Secure the middle and ring fingers with the distal strap.



2. Secure the middle strap.



3. Secure the wrist strap.



4. Finished application.

U.S. Patent No. 7278980, 7402149, 7455650, 7442177



1185 E Main St., Santa Paula, California 93060 California: 800-221-5465, National: 800-654-3241 International: 805-525-4244, fax: 805-933-2348 U.S. fax: 800-559-5975, www.hely-weber.com



Rev D 2/11