McKESSON

URGENT MEDICAL DEVICE RECALL

February 17, 2026

Dear Valued McKesson Customer:

Integra LifeSciences has notified McKesson Medical-Surgical Inc. (MMS) of an Urgent Medical Device Recall regarding all unexpired lots of their MediHoney® Wound & Burn products. This notice has been issued because during an investigation, packaging failures were identified which could lead to a breach in the sterile barrier. Affected product first shipped August 5, 2020.

This Urgent Medical Device Recall is being done with the knowledge of the U.S. Food and Drug Administration. McKesson Medical-Surgical Inc. has taken appropriate action per this notice.

For clinical inquiries, please contact Integra LifeSciences at (800) 654-2873.

A review of our records indicates that your company may have purchased items included in this notification. Carefully review the information in this letter and follow the instructions provided below.

Refer to the table for a list of affected item(s) distributed by McKesson Medical-Surgical

788894	31815	DRESSING, MEDIHONEY GEL STR 1.5OZ (12/CS)	All unexpired lots
788895	31805	DRESSING, MEDIHONEY GEL STR TU0.5OZ (10/BX 4BX/CS)	All unexpired lots
688573	31012	DRESSING, MEDI-HONEY 3/4"X12" (5/BX 4BX/CS)	All unexpired lots
683888	31022	DRESSING, MEDI-HONEY 2"X2" (10/BX 10BX/CS)	All unexpired lots
702880	31045	DRESSING, MEDI-HONEY 4"X5" (10/BX 5BX/CS)	All unexpired lots

McKesson Customer Instructions:

1.) Immediately discontinue use of any product matching the affected item(s) and lot number(s) listed in the item table. If you have no products matching the affected item(s) and lot number(s), no further action is needed.

2.) A copy of the Urgent Medical Device Recall from Integra LifeSciences has been included for reference.

3.) If you have product affected by this notice, fill out the McKesson Reply Form and return it to our Corporate Customer Service Center via email at MMSRecalls@McKesson.com or fax at (866) 871-6270. To ensure timely credit to your account and support the completion of this notice, please respond within 30 days.
- Please note: Any product returned in addition to or in lieu of affected product will be destroyed, without issuance of a credit. The affected product lot numbers are listed in the item table. Once the product is returned, credit will be issued to you.
- Please place a new order for replacement product if there is an immediate need.

4.) If you have further distributed any of the item(s) referenced in this notification, provide your accounts with a copy of this Urgent Medical Device Correction and request that they return the affected product directly to you.