



1

To open, pull the red tape to remove the sterile plastic seal. Remove the paper.

**▲ Caution:** DO NOT use if the seal or sterile paper has been broken.

**■ Note:** Clean the insertion area with an alcohol prep pad or other recommended skin preparation.



2

Remove the lid by pressing the three raised dots on either side of the lid firmly with one hand and lifting the flat tab of the lid with the other hand.

**▲ Caution:** DO NOT touch or bend the introducer needle.



3

Unhook the tubing from the notch on the insertion device, then gently unwind the tubing. Stop when you get to the end.

**▲ Caution:** Be careful when handling the insertion device to keep from dislodging the cannula housing.



4

To attach the infusion set tubing to the t:lock connector on the cartridge, twist the connector clockwise until tight, then twist another quarter of a turn to ensure a secure connection.



5

Fill your tubing according to the instructions in the Tandem Diabetes Care insulin pump user guide.

**▲ Caution:** DO NOT fill tubing while the infusion set is connected to your body.



6

When filling the tubing, hold the introducer needle pointing down and hold the tubing straight upwards at the t:lock connector. Make sure there are no gaps of air in the tubing.



7

Stop after three drops of insulin have appeared from the introducer needle.

**■ Note:** The approximate fill volumes are 10-15 units for 60 cm tubing, and 20-25 units for 110 cm tubing.



8

Pull up gently on the backing paper to remove it from the adhesive tape.

**▲ Caution:** Be careful not to pull the backing paper too hard.



9

Make sure the tubing hangs freely (not in the notch), then grip the lined ridges and pull the spring loader back until you hear a click.



10

Remove the needle guard from the introducer needle by gently twisting and pulling it off.



11

Slide the tubing into the closest notched edge.

**▲ Caution:** If the introducer needle is not straight, pain or minor injury may occur during insertion.



12

Prior to insertion, it is critical to verify that the plastic cannula is below the tip of the introducer needle.



13

Simultaneously squeeze the round indentations on both sides of the insertion device to insert.

**Note:** The triangle inside the insertion device points to the direction the tubing will face.



14

Massage the back of the inserter to secure the adhesive tape onto the skin.

**Note:** Skipping this step may result in your infusion set falling off early and/or a bent cannula.



15

Hold the white tabs of the insertion device and gently pull back to remove the introducer needle. Massage the tape onto the skin.



16

Fill cannula 0.3 units (6 mm cannulas) or 0.5 units (9 mm cannulas). Put the lid back on the insertion device before disposing.

**Caution:** ALWAYS check glucose levels 1-2 hours after inserting to make sure it's working properly.

## Disconnect



1

Place your finger in front of cannula housing, gently squeeze the sides near the center of the connector, and pull the connector needle straight out.



2

Insert the disconnect cover into the cannula housing.



1

Remove the circular protective cap from the connector needle. If needed, fill the infusion set with the needle pointing down until insulin drops from the needle.

**Caution:** DO NOT fill tubing while the infusion set is connected to your body.



2

Remove the disconnect cover. Insert the connector needle into the cannula housing. You should hear a click.

## Reconnect

This reference sheet is intended to supplement the instructions included with your AutoSoft 90 infusion set. Please refer to those instructions (available for download at [tandemdiabetes.com/support](http://tandemdiabetes.com/support)) for a complete list of warnings and precautions.

**Important Safety Information:** Caution: Federal Law (United States of America) restricts this device to sale by or on the order of a physician. These infusion sets are indicated for the subcutaneous infusion of insulin administered by the t:slim X2 insulin pump for the treatment of diabetes. These infusion sets are indicated for single use. For contraindications, warnings, precautions, and other important information, please refer to the instructions for use accompanying the infusion sets.

© 2022 Tandem Diabetes Care, Inc. All rights reserved. Tandem Diabetes Care, AutoSoft, t:lock, and t:slim X2 are either registered trademarks or trademarks of Tandem Diabetes Care, Inc. in the United States and/or other countries. All third-party marks are the property of their respective owners. ML-1009393\_A